

Iso-Analytical Limited

Report of Analysis

IA-R072- ²H and ¹³C Mineral Oil Laboratory Standard

This laboratory standard is intended to provide a sample of known isotope composition with ²H/¹H and ¹³C/¹²C isotope ratios stated in parts per thousand difference (‰) from V-SMOW and the V-PDB (Pee Dee Belemnite) isotope ratio standards, respectively. This laboratory standard is not certified, but is provided to allow routine checking of the overall quality of measurements performed by continuous-flow isotope ratio mass spectrometry, and may be used as part of a quality control program. It is not intended for use as a substitute for calibration materials or inter-comparison materials distributed by NIST, USGS or IAEA.

Analysis

The ²H/¹H isotope ratio of the laboratory standard was measured by elemental analyser continuous-flow isotope ratio mass spectrometry using NBS-22 (Mineral Oil, δ²H = -116.09 ‰ vs. V-SMOW-SLAP) as the calibration material. The ²H/¹H isotope ratio in the laboratory standard was measured eight times on three separate occasions.

The ¹³C/¹²C isotope ratio of the laboratory standard was measured by elemental analyser continuous-flow isotope ratio mass spectrometry using NBS-22 (Mineral Oil, δ¹³C = -30.03 ‰ vs. V-PDB-LSVEC) as the calibration material. The ¹³C/¹²C isotope ratio in the laboratory standard was measured six times on four separate occasions.

Isotope Abundance

The laboratory standard IA-R072 is compared to V-SMOW for the ²H/¹H isotope ratio and V-PDB for the ¹³C/¹²C isotope ratio. The isotope composition of the laboratory standard in ‰ relative to V-SMOW and V-PDB is:

Laboratory Standard	δ ² H _{V-SMOW} (‰) δ _m ± σ ₁	δ ¹³ C _{V-PDB} (‰) δ _m ± σ ₁
IA-R072	-148.61 ± 2.16	-29.27 ± 0.02

Note: $\delta_m = \sum_{i=1}^n \delta_i/n$; $\sigma_1 = \sqrt{\sum_{i=1}^n (\delta_m - \delta_i)^2/(n-1)}$; n = 24 for ²H and for ¹³C

May 3, 2019

Steven T. Brookes

Charles Belanger